Qingdao Kingon Medical Science and Technology Co., Ltd

Manufacture's FAA acceptance criteria Declaration of Conformity

Qingdao Kingon Medical Science and Technology Co., Ltd has determined the following POC devices conform to all applicable FAA acceptance criteria for POC carriage and use on board aircraft.

Device: Portable Oxygen Concentrator

Model No.: P2

Brand: Kingon

The above device has tested and passed the requirements of;

RTCA/DO-160G, Environmental Conditions and Test Procedures for Airborne Equipment.

Section 21, Category M of RTCA Document (DO) -160 testing

Test reports are available upon request to:

Qingdao Kingon Medical Science and Technology Co., Ltd

Manufacturer Address:

Room 301-302,No.15 Hancheng Road, Qingdao Free Trade Zone,Shandong, Chin, 266555 Factory Address:

24th Factory Building, NO. 252 Yanhe Road, Huangdao, Qingdao, Shandong, China, 266510

Tel: +86 532 58792324

E-mail: kingonmed@gr naifc@n精 安 医 疗 科 技 有 限 责 任 公 司

www.kingonmed.com

Qinqdao Kingon Medical Science & Technology Co., Ltd